

PDS No. 6550xx	PRODUCT DATA SHEET				Page 1 of 1
Revision 00	96 Well ELISA-Plates, PS, F-Bottom				
	Greiner Item-No. 6550xx				
Valid for Item-No.:	655001	655061	655091	655092	655093

1.	Description / Specification	
1.1	Description	PS ELISA-Plates, 96 well, solid F-bottom (flat), 655001: MICROLON 200, medium binding treatment 655061, -091, -092, -093: MICROLON 600, high binding treatment
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	Total volume: 382 µl (mathematically calculated) Working volume: 25-340 µl
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for immunology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	4.800 x g: swinging-bucket rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	4 years
2.7	Other Information	-

3.	Packaging	655001, -061	655091	655092	655093
3.1	Pieces / Bag	10	10	2	50
3.2	Pieces / Box	40	100	50 (secondary package; cardboard box) 400	50 (secondary package; cardboard box) 400
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)			
3.4	Other Information	Certificate of Quality to download			

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision	Date	Date	Date	
03 (655 09x) 05 (655 001, -061) 04 (655 091)	22 June 2026	26 June 2026	26 June 2026	
Date	Name	Name	Name	
17.01.2022	S. Kaelberer	Dr. L. Breth	Dr. S. Mühlfriedel	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.