


PDS No. 6531xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 04	60 Well Microtest Plate (Terasaki Plate), PS			
	Greiner Item-No. 6531xx			
Valid for Item-No.:	653102	653180	653190	

1.	Description / Specification	
1.1	Description	PS Microtest Plate (Terasaki Plate), 60 Well, conical well profile, flat well bottom, alphanumeric well coding, with lid 653102: Microbatch under Oil Applications, standard 653180, -190: HLA / Microtest Plate, physical surface treatment
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	10 µl
1.4	Material / Resin	Plate and lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Plate and lid: clear
1.6	Sterilization	No
1.7	Quality Control	- Raw Material-Control: physical and immunological testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	653102: General laboratory product for the processing and storage of samples to be used by qualified personnel in a laboratory environment. 653180: General laboratory product for cell culture to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	653102: n/a 653180, -190: 1 year
2.7	Other Information	-

3.	Packaging	653102	653180	653190
3.1	Pieces / Bag	10 / bag	10 / bag	120 / folding carton
3.2	Pieces / Box	580	270	480
3.3	Lot-No.	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)		
3.4	Other Information	-		

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 9 February 2022	Date 27 May 2022	Date 30 May 2022	
Date 28.11.2014	Name S. Kaelberer	Name P. Wachter	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.