


PDS No. 67605x	PRODUCT DATA SHEET					Page 1 of 1
Revision 05	Sealer BREATHseal™					 greiner bio-one
	Greiner Item-No. 67605x					
Valid for Item-No.:	676050	676051 (sterile)				

1.	Description / Specification	
1.1	Description	Gas-permeable, piercable membrane coated with acrylate adhesive, consisting of heat sealed rayon fibres
1.2	Dimensions	153 x 83 mm Thickness: 0.14 mm
1.3	Volume	-
1.4	Material / Resin	<u>Foil</u> : Rayon-fibres, porous, non woven tape <u>Glue</u> : acrylate adhesive <u>Paper</u> : Poly-coated Kraft-paper with silicone
1.5	Colour	Foil: white, translucent
1.6	Sterilization	676050: no 676051: SAL 10 ⁻³
1.7	Quality Control	-
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	<ul style="list-style-type: none"> - Minimal skin irritation (Acute Primary Skin Irritation). No evidence of induced contact sensitization - In Vitro Cytotoxicity (Agar Overlay) - Acute Primary Skin Irritation in Albino Rabbits - Repeat Skin Irritation in Albino Rabbits - Repeat Insult Patch Test (Draize) in Humans <p>These tests are in accordance with the ISO10993 Part-1 "Biological Evaluation of Medical Devices", as put forth by the FDA</p> <ul style="list-style-type: none"> - MVTR (Moisture Vapour Transmission Rate): 4200g/m²/24h
2.2.	Temperature range	N/A
2.3	Autoclavability	N/A
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	N/A
2.6	Shelf life	2 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	50
3.2	Pieces / Box	500
3.3	Lot-No.	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 21 April 2015	Date 22 April 2015	Date 22 April 2015	
Date 25.07.2012	Name S. Kaelberer	Name Dr. U. Honisch	Name A. Schulz	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.