


PDS No. 76207x	PRODUCT DATA SHEET				Page 1 of 1
Revision 07	96 Well Strip Plate, PS, F-Bottom				
	Greiner Item-No. 76207x				
Valid for Item-No.:	762070	762071	762074	762075	762076
	762077				

1.	Description / Specification	
1.1	Description	PS Strip plate, 12 x F8 strips mounted in frame, solid F-bottom (flat) 762070: MICROLON 200, medium binding 762071: MICROLON 600, high binding 762074: LUMITRAC 600, high binding 762075: LUMITRAC 200, medium binding 762076: FLUOTRAC 200, medium binding 762077: FLUOTRAC 600, high binding
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	Total volume: 388 µl Working volume: 20-350 µl
1.4	Material / Resin	Strips and Frame: PS (Polystyrene), free of heavy metal
1.5	Colour	Frame: white 762070, -071: strip: clear 762074, -075: strip: white 762076, -077: strip: black
1.6	Sterilization	No
1.7	Quality Control	- Raw Material-Control: physical and immunological testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for immunology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	4 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	100
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 06	Date 21 December 2021	Date 17 January 2022	Date 17 January 2022	
Date 03.12.2014	Name S. Kaelberer	Name R. Daum	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.