



PDS No.9765xx	<b>PRODUCT DATA SHEET</b>			Page 1 of 2
Revision 04	Rack with 96 Cryo.s™ Biobanking Tubes 300 µl (Datamatrix Coded)			
	Greiner Item-No. 9765xx			
Valid for Item-No.:	<b>97656x</b> (sterile)	<b>976570</b>	<b>97658x</b>	

1.	Description / Specification	
1.1	Description	<p>Rack with 300 µl Cryo.s™ Biobanking Tubes, stackable, with a footprint that corresponds to the microplate standard of the American National Standards Institute (ANSI-Standard), capacity for 96 Cryo.s™ Biobanking Tubes 300 µl, rack bottom with scanning windows for the Datamatrix, rotation stoppers, snap-in lid.</p> <p><u>Biobanking Tubes 300 µl</u> with screw caps (caps have a silicone gasket and internal thread, different colours available), unique 2D codes on bottom, without writing area and graduation</p> <p><u>Manual decapping tool</u> for manual decapping, capping and contamination-free storage of screw-cap, contained in each box</p> <p>- 2D Code on tube (G8xxxxxxx) - 2D Code and linear barcode 128 on rack (G9xxxxxxx)</p> <p>All pre-produced 2D codes on tubes and racks are featured with a symbol size of 14x14</p> <p>976561, -563, -564, -565, -566, -567, -568, -569: rack with 96 capped 300 µl Cryo.s™, sterile 976570: rack with 96 uncapped 300 µl Cryo.s™, non-sterile 976580, -583, -584, -585, -586, -587, -588, -589: rack with 96 capped 300 µl Cryo.s™, non-sterile</p> <p><i>Customised barcoding option (9765xx-2D1) available via order form (customised sequences, Cat.-No. F071003)</i></p>
1.2	Dimensions	See Customer Drawing
1.3	Volume	Tube / working volume: 235 µl
1.4	Material / Resin	<p><u>Tube / cap</u>: PP (Polypropylene), medical grade and USP class VI certified <u>Rack / lid</u>: PC (Polycarbonate) <u>Gasket</u>: Silicone, medical grade <u>Manual capping tool</u>: PS (Polystyrene)</p>
1.5	Colour	<p><u>Tube</u>: natural <u>Cap</u>: 976561, -580: natural 976563, -583: red 976564, -584: blue 976565, -585: yellow 976566, -586: green 976567, -587: black 976568, -588: pink 976569, -589: brown</p> <p><u>Gasket (cap)</u>: blue <u>Rack</u>: black <u>Lid</u>: transparent <u>Manual capping tool</u>: blue</p>
1.6	Sterilization	976561, -563, -564, -565, -566, -567, -568, -569: SAL 10 <sup>-6</sup> 976570, -580, -583, -584, -585, -586, -587, -588, -589: no
1.7	Quality Control	<p>- <u>Raw Material-Control (tubes)</u>: physical testing</p> <p>- <u>Product-Control</u>: testing of attributive and variable characteristics in accordance with the valid specification</p>

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 4 December 2015	Date 2 February 2016	Date 3 February 2016	
Date 04.05.2015	Name S. Kaelberer	Name Dr. S. Mühlfriedel	Name Dr. A. Ganser	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.

PDS No.9765xx	<b>PRODUCT DATA SHEET</b>			Page 2 of 2
Revision 04	Rack with 96 Cryo.s™ Biobanking Tubes 300 µl (Datamatrix Coded)			
	Greiner Item-No. 9765xx			
Valid for Item-No.:	<b>97656x</b> (sterile)	<b>976570</b>	<b>97658x</b>	

1.8	Other Information	<p>Cryo.s™ with datamatrix and accessory storage racks are optimised for automated laboratory equipment such as decappers, liquid handling systems and automated storage systems. Greiner Bio-One recommends closing 96 way Cryo.s™ with the automated decapper from Hamilton Bonaduz AG or an equivalent device at 7 Ncm. For manual contamination-free application and removal of screw caps and for manual picking of individual tubes from rack, a manual tool is available.</p> <p>A list of compatible automated equipment may be requested from GBO.</p> <p>For single use only</p>
-----	-------------------	--

<b>2.</b>	<b>Features</b>	
2.1	Basic features	Tubes: free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic
2.2	Temperature range	-196°C to 121°C
2.3	Autoclavability	N/A
2.4	Centrifugation, max. RCF	3.000 x g: swinging-bucket rotor (with slip-resistant mat)
2.5	Chemical Resistance	<p>Tube / material: See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a></p> <p>Tube / datamatrix: resistant to → acetic acid (1 %), hydrochloric acid (25 %), sulphuric acid (4.9%), ethanol, methanol, isopropanol, DMSO</p> <p>Rack / lid: resistant to → acetic acid (1 %), methanol, ethanol (96%), isopropanol, sulphuric acid (4.9%), DMSO</p>
2.6	Shelf life	<p>976561, -563, -564, -565, -566, -567, -568, -569: 5 years after month of production</p> <p>976570, -580, -583, -584, -585, -586, -587, -588, -589: n/a</p>
2.7	Other Information	Recommended for storage in the freezer or above liquid nitrogen. Must not be used in liquid phase of nitrogen!

<b>3.</b>	<b>Packaging</b>	<b>976561, -570, -580</b>	<b>976563, -564, -565, -566, -567, -568, -569, -583, -584, -585, -586, -587, -588, -589</b>
3.1	Pieces / Rack	96	96
3.2	Pieces / Box	960 tubes, 10 racks, 1 manual capping tool	480 tubes, 5 racks, 1 manual capping tool
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality	Certificate of Quality

<b>4.</b>	<b>Other Information</b>
4.1	Research use only. Not for diagnostics.

Data Sheet subject to change without notice!

<b>Prior Issue</b>	<b>Drawn</b>	<b>Approved</b>	<b>Released</b>	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 4 December 2015	Date 2 February 2016	Date 3 February 2016	
Date 04.05.2015	Name S. Kaelberer	Name Dr. S. Mühlfriedel	Name Dr. A. Ganser	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.