


PDS No. 6712xx	<b>PRODUCT DATA SHEET</b>				Page 1 of 1
Revision 06	Sapphire PCR Tubes, 0.2 ml, Domed Cap				 greiner bio-one
	Greiner Item-No. 6712xx				
Valid for Item-No.:	<b>671201</b>	<b>671273</b>	<b>671274</b>	<b>671275</b>	<b>671276</b>
	<b>671277</b>	<b>671281</b>			

1.	Description / Specification	
1.1	Description	Sapphire PCR tubes, 0.2 ml with attached domed cap. Frosted tube caps for easy labelling.
1.2	Dimensions	See Customer Drawing
1.3	Volume	0,2 ml
1.4	Material / Resin	PP (Polypropylene), free of heavy metal
1.5	Colour	671201: clear 671273: red 671274: blue 671275: green 671276: yellow 671277: violet 671281: assorted (red, blue, green, yellow, violet)
1.6	Sterilization	No
1.7	Quality Control	<u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens
2.2	Temperature range	-20°C to +105°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	5 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Box	1.000
3.2	Pieces / Case	10.000
3.3	Lot-No.	5-digit 12345 (digit 1 & 2 identify the year and 3 & 4 identify the week)
3.4	Other Information	Packaging labelled with product symbols for easier identification and warehousing

4.	Other Information	
		-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 05	Date 13 April 2016	Date 30 June 2016	Date 1 July 2016	
Date 22.04.2015	Name S. Kaelberer	Name Dr. R. Heller	Name Dr. A. Ganser	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.