


PDS No. 5410xx	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 05	Cell Scrapers			
	Item-No. 5410xx			
Valid for Item-No.:	<b>541070</b> (sterile)	<b>541080</b> (sterile)		

1.	Description / Specification	
1.1	Description	Cell scraper, pivoted scraper blade, sterile 541070: short handle 541080: long handle
1.2	Dimensions	See Customer Drawing
1.3	Volume	-
1.4	Material / Resin	Handle: PS (Polystyrene), free of heavy metal Blade: HDPE (High Density Polyethylene), free of heavy metal
1.5	Colour	Handle: blue Blade: white
1.6	Sterilization	SAL 10 <sup>-3</sup>
1.7	Quality Control	Raw Material-Control: physical testing Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for mechanical removal of adherent cells to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	-
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	1
3.2	Pieces / Box	100
3.3	Lot-No.	F YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information
4.1	Research use only. Not for diagnostics.

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 2 February 2023	Date 17 April 2023	Date 18 April 2023	
Date 20.04.2015	Name S. Kaelberer	Name Dr. R. Daum	Name Dr. Chen-Ket Chai	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.