


PDS No. 6500xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 06	96 Well ELISA Microplate, PS, MICROLON, U-bottom			
	Greiner Item-No. 6500xx			
Valid for Item-No.:	650001	650061		

1.	Description / Specification	
1.1	Description	PS Microplate, 96 well, clear, solid U-bottom, alphanumeric well coding 650001: MICROLON 200, medium binding 650061: MICROLON 600, high binding
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	Total volume: 323 µl (mathematically calculated) Working volume: 40 - 280 µl
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for immunology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	- For single use only - MICROLON logo on plate

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	1.000 x g: swinging-bucket rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	4 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	40
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 05	Date 20 December 2021	Date 17 January 2022	Date 17 January 2022	
Date 27.11.2014	Name S. Kaelberer	Name R. Daum	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.