PDS No. 1222xx, 1222xx-2DG, 1222xx-2D3		PRODUCT	DATA SHEET	-	Page 1 of 2
Revision 13	Cryo.s, 2 ml, Round Bottom, Starfoot Base UK CF ○197				greiner
	Item-No. 1222xx, 1222xx-2DG, 1222xx-2D3		BIO-ONE		
Valid for Item-No.:	122261 (sterile)	122263 (sterile)	122272 (sterile)	122277 (sterile)	122278 (sterile)
	122279 (sterile)	122280 (sterile)	1222xx-2DG	1222xx-2D3	
			(sterile)	(sterile)	

1.	Description / Specification		
1.1	Description	Cryo.s, 2 ml, with different coloured screw caps (caps have a silicone gasket), internal thread, round bottom, starfoot base and additional cap inserts for labelling, sterile. 122261, -289: without writing area	
		122263, -272, -277, -278, -279, -280: with writing area and graduation	
		122263, -277,-278, -279, -280-2DG: with linear barcode 128, human readable ID, graduation and ECC200 Datamatrix-Code, symbol size 14x14, on tube bottom, code sequence 'G200000001', 'G200000002' etc., for screv cap colour see 1.5	
		122263, -277,-278, -279, -280-2D3: with linear barcode 128, human readable ID, graduation on tube side and ECC200 Datamatrix-Code, symbol size 14x14, on tube bottom, code sequence customized as specified in order form CatNo. F071005, for screw cap colour see 1.5	
		Additional inserts, CatNo. 304 134 (50 pieces per bag)	
1.2	Dimensions	See Customer Drawing	
1.3	Volume	Working volume: up to 2.0 ml	
1.4	Material / Resin	Tube: PP (Polypropylene), free of heavy metal Screw cap: PP (Polypropylene), free of heavy metal Gasket: Silicone Small cap insert: PS (Polystyrene), free of heavy metal	
4.5		1222xx-2DG, 1222xx-2D3: Datamatrix disc: PP (Polypropylene)	
1.5	Colour	Tube: natural graduation: 122263, -272, -277, -278, -279, -280: white Screw cap: 122261, -263: natural 122272: orange 122277: green 122278: yellow 122279: blue 122280: red 1222xx-2D3, 1222xx-2DG: xx = colour code of screw cap with xx = 63 for natural screw cap, xx = 77 for green screw cap, xx = 78 for yellow screw cap, xx = 79 for blue screw cap, xx = 80 for red screw cap) Gasket: translucent Small caps insert: white 1222xx-2DG, 1222xx-2D3: white Datamatrix at the bottom and linear barcode on side	
1.6	Sterilization	SAL 10 ⁻⁶	
1.7	Quality Control	Raw Material-Control: physical testing Product-Control: testing of attributive and variable characteristics in accordance with the valid specification	
1.8	Intended Use	For storage of tissue, cells, fungi, bacteria, spores, cellular extracts or body fluids at ultra-low temperature for research and development as well as diagnostic purposes. Tubes must not be stored in the liquid phase of liquid nitrogen, but only in the gas phase above. Tubes are not intended for any	

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date		document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This
12	3 April 2023	3 May 2023	14 September 2023	document may not be reproduced for any
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,
10.02.2023	S. Kaelberer	Dr. S. Mühlfriedel	Dr. Chen-Ket Chai	and copyright are reserved.

PDS No. 1222xx, 1222xx-2DG, 1222xx-2D3	PRODUCT DATA SHEET			Page 2 of 2	
Revision 13	Cryo.s, 2 ml, Round Bottom, Starfoot Base UK CF €0197			greiner	
Iter		-No. 1222xx, 1222xx-2DG, 1222xx-2D3			BIO-ONE
Valid for Item-No.:	122261 (sterile)	122263 (sterile)	122272 (sterile)	122277 (sterile)	122278 (sterile)
	122279 (sterile)	122280 (sterile)	1222xx-2DG	1222xx-2D3	
			(sterile)	(sterile)	

		application in the context of reproductive medicine.	
1.9	Other Information	For single use only	

2.	Features		
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic	
2.2	Temperature range	For applications such as sample transportation, storage, thawing, processing and centrifugation: -196°C to +37°C	
2.3	Autoclavability	No	
2.4	Centrifugation, max. RCF	5.800 x g: swinging-bucket rotor 34.000 x g: fixed-angle rotor	
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html	
2.6	Shelf life	5 years	
2.7	Other Information	To achieve tight tube closure (liquid barrier) tubes must be closed with torque of 20.5 Ncm	

3.	Packaging	
3.1	Pieces / Bag	100
3.2	Pieces / Box	500
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Additional in each box: Safety precaution
		3 bags with 50 cap inserts per bag (150 pieces)
		Certificate of Quality to download

4.	Other Information
4.1	Laboratory use only

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date	Date	document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This
12	3 April 2023	3 May 2023	14 September 2023	document may not be reproduced for any
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,
10.02.2023	S. Kaelberer	Dr. S. Mühlfriedel	Dr. Chen-Ket Chai	and copyright are reserved.