


PDS No. 73118x	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 04	Disposable Inoculation Needle, PS			
	Greiner Item-No. 73118x			
Valid for Item-No.:	<b>731180</b> (sterile)	<b>731181</b> (sterile)	<b>731185</b> (sterile)	

1.	Description / Specification	
1.1	Description	PS disposable inoculation needle
1.2	Dimensions	See Customer Drawing
1.3	Volume	-
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Yellow
1.6	Sterilization	SAL 10 <sup>-3</sup>
1.7	Quality Control	- <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for bacteriology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Hydrophobic
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	731180	731181	731185
3.1	Pieces / Bag	50	1	10
3.2	Pieces / Box	2.000	600	3.000
3.3	Lot-No.	F YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	F YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	F YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-	-	

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 2 February 2023	Date 13 April 2023	Date 13 April 2023	
Date 22.04.2015	Name S. Kaelberer	Name Dr. R. Daum	Name Dr. Chen-Ket Chai	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.