


PDS No.97856x, -561-2D1	PRODUCT DATA SHEET			Page 1 of 2
Revision 08	Rack with 96 Cryo.s Biobanking Tubes 1000 µl (Datamatrix Coded), Sterile		UK CA	CE ₀₁₉₇ 
Greiner Item-No. 97856x / 978561-2D1				
Valid for Item-No.:	97856x (sterile)	978561-2D1 (sterile)		

1.	Description / Specification	
1.1	Description	<p><u>Rack with 1000 µl Cryo.s Biobanking Tubes</u>, stackable, with a footprint that corresponds to the microplate standard of the American National Standards Institute (ANSI-Standard), capacity for 96 Cryo.s Biobanking Tubes 1000 µl, rack bottom with scanning windows for the Datamatrix, rotation stoppers, snap-in lid.</p> <p><u>Biobanking Tubes 1000 µl</u> with screw caps (caps have a silicone gasket and internal thread, different colours available), unique 2D codes on bottom, without writing area and graduation</p> <p><u>Manual decapping tool</u> for manual decapping, capping and contamination-free storage of screw cap, contained in each box</p> <p>- 2D Code on tube (G8xxxxxxx) - 2D Code and linear barcode 128 on rack (G9xxxxxxx)</p> <p>All pre-produced 2D codes on tubes and racks are featured with a symbol size of 14x14</p> <p>978561, -563, -564, -565, -566, -567, -568, -569: rack with 96 capped 1000 µl Cryo.s, sterile 978561-2D1: customised barcoding option available via order form (customised sequences, Cat. - No. F071003)</p>
1.2	Dimensions	See Customer Drawing
1.3	Volume	Tube / working volume: 975 µl
1.4	Material / Resin	<p><u>Tube / cap</u>: PP (Polypropylene), medical grade and USP class VI certified</p> <p><u>Rack / lid</u>: PC (Polycarbonate),</p> <p><u>Gasket</u>: Silicone, medical grade</p> <p><u>Manual capping tool</u>: PS (Polystyrene)</p>
1.5	Colour	<p><u>Tube</u>: natural</p> <p><u>Cap</u>: 978561: natural 978563: red 978564: blue 978565: yellow 978566: green 978567: black 978568: pink 978569: brown</p> <p><u>Gasket (cap)</u>: blue</p> <p><u>Rack</u>: black</p> <p><u>Lid</u>: transparent</p> <p><u>Manual capping tool</u>: blue</p>
1.6	Sterilization	SAL 10 ⁻⁶
1.7	Quality Control	<p>- <u>Raw Material-Control (tubes)</u>: physical testing</p> <p>- <u>Product-Control</u>: testing of attributive and variable characteristics in accordance with the valid specification</p>

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 07	Date 6 June 2024	Date 5 August 2024	Date 7 August 2024	
Date 13.09.2023	Name S. Kaelberer	Name Dr. S. Mühlfriedel	Name Dr. C.-K. Chai	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.

PDS No.97856x, -561-2D1	PRODUCT DATA SHEET			Page 2 of 2
Revision 08	Rack with 96 Cryo.s Biobanking Tubes 1000 µl (Datamatrix Coded), Sterile		UK CA	CE ⁰¹⁹⁷
	Greiner Item-No. 97856x / 978561-2D1			
Valid for Item-No.:	97856x (sterile)	978561-2D1 (sterile)		



1.8	Intended Use	For storage of tissue, cells, fungi, bacteria, spores, cellular extracts or body fluids at ultra-low temperature for research and development as well as diagnostic purposes*. Tubes must not be stored in the liquid phase of liquid nitrogen, but only in the gas phase above. Tubes are not intended for any application in the context of reproductive medicine. *Intended use for USA: Research use only
1.9	Other Information	Cryo.s with datamatrix and accessory storage racks are optimised for automated laboratory equipment such as decappers, liquid handling systems and automated storage systems. Greiner Bio-One recommends closing 96-way Cryo.s with the automated decapper from Hamilton Bonaduz AG or an equivalent device at 7 Ncm. For manual contamination-free application and removal of screw caps and for manual picking of individual tubes from rack, a manual tool is available. A list of compatible automated equipment may be requested from GBO. For single use only

2.	Features	
2.1	Basic features	<u>Tubes</u> : free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic
2.2	Temperature range	For application: -196°C to 37°C
2.3	Autoclavability	N/A
2.4	Centrifugation, max. RCF	3.000 x g: swinging-bucket rotor (with slip-resistant mat)
2.5	Chemical Resistance	<u>Tube / material</u> : See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html <u>Tube / datamatrix</u> : resistant to → acetic acid (1 %), hydrochloric acid (25 %), sulphuric acid (4.9%), ethanol, methanol, isopropanol, DMSO <u>Rack / lid</u> : resistant to → acetic acid (1 %), methanol, ethanol (96%), isopropanol, sulphuric acid (4.9%), DMSO
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	978561, 978561-2D1	978563, -564, -565, -566, -567, -568, -569
3.1	Pieces / Rack	96	96
3.2	Pieces / Box	960 tubes, 10 racks, 1 manual capping tool	480 tubes, 5 racks, 1 manual capping tool
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download	Certificate of Quality to download

4.	Other Information
4.1	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 07	Date 6 June 2024	Date 5 August 2024	Date 7 August 2024	
Date 13.09.2023	Name S. Kaelberer	Name Dr. S. Mühlfriedel	Name Dr. C.-K. Chai	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.