


PDS No. 6832xx	<b>PRODUCT DATA SHEET</b>				Page 1 of 1
Revision 07	Sapphire PCR Tubes, 0.2 ml, Flat Cap				
	Greiner Item-No. 6832xx				
Valid for Item-No.:	<b>683201</b>	<b>683271</b>	<b>683273</b>	<b>683274</b>	<b>683275</b>
	<b>683276</b>	<b>683277</b>			

1.	Description / Specification	
1.1	Description	Sapphire PCR tubes, 0.2 ml with attached flat cap
1.2	Dimensions	See Customer Drawing
1.3	Volume	0.2 ml
1.4	Material / Resin	PP (Polypropylene), free of heavy metal
1.5	Colour	683201: clear 683271: assorted (red, blue, green, yellow, violet) 683273: red 683274: blue 683275: green 683276: yellow 683277: violet
1.6	Sterilization	No
1.7	Quality Control	Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for molecular biology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens
2.2	Temperature range	For application: -20°C to +105°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Box	1.000
3.2	Pieces / Case	10.000
3.3	Lot-No.	5-digit 12345 (digit 1 & 2 identify the year and 3 & 4 identify the week)
3.4	Other Information	Packaging labelled with product symbols for easier identification and warehousing

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 06	Date 20 August 2021	Date 16 November 2021	Date 17 November 2021	
Date 01.07.2016	Name S. Kaelberer	Name Dr. M. Langbein	Name A. Illig	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.