


PDS No. 781856	PRODUCT DATA SHEET		Page 1 of 1
Revision 09	384 Well SensoPlate Plus, PS, Glass Bottom		
	Greiner Item-No. 781856		
Valid for Item-No.:	781856		

1.	Description / Specification	
1.1	Description	PS Microplate, 384 well, F-glass bottom (flat), rounded square well design, alphanumeric well coding, optimised microplate geometry, low curvature, without lid, non-sterile
1.2	Dimensions	Plate: see Customer Drawing Glass bottom: 170 µm (+/- 20 µm)
1.3	Volume per well	Total volume: 133 µl (mathematically calculated) Working volume: 10 – 130 µl
1.4	Material / Resin	Plate: PS (Polystyrene), free of heavy metal Glass bottom: clear borosilicate
1.5	Colour	Plate: black Glass bottom: clear
1.6	Sterilization	No
1.7	Quality Control	Raw Material-Control: physical testing Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for the processing and storage of samples to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Adhesive: resistant to DMSO and 70% ethanol, compatible with cell culture media, low autofluorescence
2.2	Temperature range	For application: + 4°C to +37°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Shelf life	2 years
2.6	Other Information	-

3.	Packaging	
3.1	Pieces / Air Cushion Bag	1
3.2	Pieces / Box	16
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 08	Date 1 September 2022	Date 16 September 2022	Date 21 September 2022	
Date 10.09.2018	Name S. Kaelberer	Name P. Wachter	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.