


PDS No. 655163	<b>PRODUCT DATA SHEET</b>		Page 1 of 1
Revision 02	96 Well Microplate, PS, F-Bottom, Sterile		
	Greiner Item-No. 655163		
Valid for Item-No.:	<b>655163</b> (sterile)		

1.	Description / Specification	
1.1	Description	PS Microplate, 96 well, clear, solid F-bottom (flat), no chimney well, sterile
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	Total volume: 382 µl (mathematically calculated) Working volume: 25-340 µl
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	clear
1.6	Sterilization	SAL 10 <sup>-3</sup>
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for the processing and storage of samples to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	4800 x g: swinging-bucket rotor
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	25
3.2	Pieces / Box	375
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 01	Date 9 February 2022	Date 27 May 2022	Date 30 May 2022	
Date 16.05.2017	Name S. Kaelberer	Name P. Wachter	Name A. Illig	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.