


PDS No. 160101	<b>PRODUCT DATA SHEET</b>		Page 1 of 1
Revision 04	Tube, 12 ml, PS, Round Bottom		
	Item-No. 160101		
Valid for Item-No.:	<b>160101</b>		

1.	Description / Specification	
1.1	Description	PS tube, round bottom, non-sterile  Appropriate grip stopper, Cat.-No. 310321 (translucent)
1.2	Dimensions	See Customer Drawing
1.3	Volume	Working volume: 10 ml
1.4	Material / Resin	<u>Tube</u> : PS (Polystyrene), free of heavy metal <u>Grip stopper</u> : LDPE (Low Density Polyethylene), free of heavy metal
1.5	Colour	clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for the processing, centrifugation, and storage of samples to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	3.500 x g: swinging-bucket rotor 5.000 x g: fixed-angle rotor
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	N/A
2.7	Other Information	-

3.	Packaging	160101
3.1	Pieces / Bag	1.600
3.2	Pieces / Box	1.600
3.3	Lot-No.	F YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 10 November 2023	Date 14 November 2023	Date 15 November 2023	
Date 28.02.2023	Name S. Kaelberer	Name Dr. T. Schreiber	Name Dr. Chen-Ket Chai	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.