


PDS No. 6576xx	PRODUCT DATA SHEET				Page 1 of 1
Revision 06	ThinCert® Cell Culture Inserts for 6 Well Plates				
	Greiner Item-No. 6576xx				
Valid for Item-No.:	657610 (sterile)	657630 (sterile)	657631 (sterile)	657638 (sterile)	657640 (sterile)
	657641 (sterile)				

1.	Description / Specification																						
1.1	Description	ThinCert® Cell Culture Inserts, hanging geometry, sealed capillary pore membrane, physical surface treatment, sterile																					
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 33%;">Pore size (µm)</th> <th style="width: 33%;">Membrane thickness (µm)</th> <th style="width: 33%;">Pore density (cm²)</th> </tr> </thead> <tbody> <tr> <td>657610: 1,0</td> <td>22 +/- 3</td> <td>657610: 2 x 10⁶</td> </tr> <tr> <td>657630: 3,0</td> <td>20 +/- 3</td> <td>657630: 0,6 x 10⁶</td> </tr> <tr> <td>657631: 3,0</td> <td>20 +/- 3</td> <td>657631: 2 x 10⁶</td> </tr> <tr> <td>657638: 8,0</td> <td>15 +/- 3</td> <td>657638: 0,15 x 10⁶</td> </tr> <tr> <td>657640: 0,4</td> <td>22 +/- 3</td> <td>657640: 1 x 10⁸</td> </tr> <tr> <td>657641: 0,4</td> <td>22 +/- 3</td> <td>657641: 2 x 10⁶</td> </tr> </tbody> </table>	Pore size (µm)	Membrane thickness (µm)	Pore density (cm ²)	657610: 1,0	22 +/- 3	657610: 2 x 10 ⁶	657630: 3,0	20 +/- 3	657630: 0,6 x 10 ⁶	657631: 3,0	20 +/- 3	657631: 2 x 10 ⁶	657638: 8,0	15 +/- 3	657638: 0,15 x 10 ⁶	657640: 0,4	22 +/- 3	657640: 1 x 10 ⁸	657641: 0,4	22 +/- 3	657641: 2 x 10 ⁶
Pore size (µm)	Membrane thickness (µm)	Pore density (cm ²)																					
657610: 1,0	22 +/- 3	657610: 2 x 10 ⁶																					
657630: 3,0	20 +/- 3	657630: 0,6 x 10 ⁶																					
657631: 3,0	20 +/- 3	657631: 2 x 10 ⁶																					
657638: 8,0	15 +/- 3	657638: 0,15 x 10 ⁶																					
657640: 0,4	22 +/- 3	657640: 1 x 10 ⁸																					
657641: 0,4	22 +/- 3	657641: 2 x 10 ⁶																					
1.2	Dimensions	See Customer Drawing Culture surface: 452,4 mm ²																					
1.3	Volume	Working volume ThinCert®: 1,0 – 3,6 ml Working volume well: 2,0 – 4,15 ml																					
1.4	Material / Resin	ThinCert® inserts: PS (Polystyrene), free of heavy metal Capillary pore membrane: PET, free of heavy metal																					
1.5	Colour	ThinCert® inserts: clear Capillary pore membrane: 657610, -630, -641: transparent 657631, -638, -640: translucent																					
1.6	Sterilisation	SAL 10 ⁻³																					
1.7	Quality Control	Product-Control: testing of attributive and variable characteristics in accordance with the valid specification																					
1.8	Intended Use	General laboratory product for cell culture to be used by qualified personnel in a laboratory environment																					
1.9	Other Information	USP class VI certified																					

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	4 years
2.7	Other Information	-

3.	Packaging	ThinCert® inserts	Multiwell plates
3.1	Pieces (single packed; paper/plastic)	1	1
3.2	Pieces / Box	24	4
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	
3.4	Other Information	Certificate of Quality to download	

4.	Other Information
4.1	Research use only. Not for diagnostics.

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 05	Date 9 August 2023	Date 10 August 2023	Date 10 August 2023	
Date 23.06.2021	Name S. Kaelberer	Name Dr. T. Binder	Name Dr. C.-K. Chai	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.