


PDS No. 6291xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 09	Contact Dish, 65 x 15 mm, Sterile			
	Greiner Item-No. 6291xx			
Valid for Item-No.:	629161 (sterile)	629180 (sterile)		

1.	Description / Specification	
1.1	Description	Contact dish with square divisions, graduation, sterile. The special design allows direct contact of media and surface for hygiene monitoring. The germs that might be present stick to the media and colony units can be calculated due to the 10 x 10 graduation. 629161: no vents 629180: with vents
1.2	Dimensions	See Customer Drawings
1.3	Volume	-
1.4	Material / Resin	Dish: PS (Polystyrene), free from heavy metal Lid: PS (Polystyrene), free from heavy metal
1.5	Colour	Dish: clear Lid: clear
1.6	Sterilisation	SAL 10 ⁻³
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for bacteriology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	20
3.2	Pieces / Box	600
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 08	Date 5 September 2022	Date 14 September 2022	Date 14 September 2022	
Date 03.03.2015	Name S. Kaelberer	Name Dr. R. Daum	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.