


| | | | | |
|---------------------|--------------------------------|---------------|-------------------------|---|
| PDS No. 7802xx | PRODUCT DATA SHEET | | | Page 1 of 1 |
| Revision 07 | 96 Well MASTERBLOCK®, PP, 1 ml | | |  |
| | Greiner Item-No. 7802xx | | | |
| Valid for Item-No.: | 780201 | 780215 | 780261 (sterile) | |

| 1. | Description / Specification | |
|-----|-----------------------------|---|
| 1.1 | Description | PP Masterblock®, 96 well, 1ml, solid U-bottom, alphanumeric well coding 780201, -215: non-sterile 780261: sterile Appropriate CapMats, Cat.-No. 381070, 381061 |
| 1.2 | Dimensions | See Customer Drawing |
| 1.3 | Volume per well | Total volume: 1.22 ml (mathematically calculated) Working volume: 0.05 – 1.1 ml (at room temp.) 0.05 – 1.0 ml (at -20°C) |
| 1.4 | Material / Resin | PP (Polypropylene), free of heavy metal |
| 1.5 | Colour | Translucent |
| 1.6 | Sterilization | 780201, -215: no 780261: SAL 10 ⁻³ |
| 1.7 | Quality Control | - Raw Material-Control: physical testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification |
| 1.8 | Intended Use | General laboratory product for the processing and storage of samples to be used by qualified personnel in a laboratory environment. |
| 1.9 | Other Information | For single use only |

| 2. | Features | |
|-----|--------------------------|--|
| 2.1 | Basic features | Free of detectable DNase/RNase, human DNA and pyrogens |
| 2.2 | Temperature range | For application: -196°C to +121°C |
| 2.3 | Autoclavability | No |
| 2.4 | Centrifugation, max. RCF | 4.800 x g: swinging-bucket rotor |
| 2.5 | Chemical Resistance | See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html |
| 2.6 | Shelf life | 780201, -215: n/a 780261: 5 years |
| 2.7 | Other Information | - |

| 3. | Packaging | 780201, -261 | 780215 |
|-----|-------------------|--|--------|
| 3.1 | Pieces / Bag | 1 | 5 |
| 3.2 | Pieces / Box | 50 | 50 |
| 3.3 | Lot-No. | E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.) | |
| 3.4 | Other Information | Certificate of Quality to download | |

| 4. | Other Information |
|----|-------------------|
| | - |

Data Sheet subject to change without notice!

| Prior Issue | Drawn | Approved | Released | CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved. |
|--------------------|------------------------|---------------------------|---------------------------|---|
| Revision 06 | Date 30 August 2022 | Date 16 September 2022 | Date 21 September 2022 | |
| Date 23.03.2016 | Name S. Kaelberer | Name P. Wachter | Name A. Illig | |

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.