


PDS No. 18917x	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 05	Tube, 7 ml, PS, Screw Cap			
	Greiner Item-No. 18917x			
Valid for Item-No.:	<b>189170</b>	<b>189171</b> (sterile)	<b>189175</b>	

1.	Description / Specification	
1.1	Description	PS Tube with flat bottom, 7 ml, Screw Cap 189170: no Label, non-sterile 189171: no Label, sterile 189175: plain Label, non-sterile
1.2	Dimensions	See Customer Drawing
1.3	Volume	7 ml
1.4	Material / Resin	Tube: PS (Polystyrene), free of heavy metal Cap: PP (Polypropylene), free of heavy metal
1.5	Colour	Tube: clear Cap: white 189175: white Label
1.6	Sterilization	189170, -175: no 189171: yes (Ethylene oxide method / EO)
1.7	Quality Control	- Raw Material-Control: physical testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for the processing, centrifugation and storage of samples to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation	7.200 x g: fixed-angle rotor
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	189170, -175: n/a 189171: 5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	700
3.2	Pieces / Box	700
3.3	Lot-No.	YYWWXX (year, week, day, shift)
3.4	Other Information	-

4.	Other Information

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 18 August 2021	Date 23.09.2021	Date 05.11.2021	
Date 20.01.2015	Name S. Kaelberer	Name Dr. T. Schreiber	Name A. Illig	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.