


PDS No. 675001/ 675061	PRODUCT DATA SHEET			Page 1 of 1
Revision 04	96 Well ELISA Microplate, PS, MICROLON, Solid Bottom, Half Area			
	Item-No. 675001 / 675061			
Valid for Item-No.:	675001	675061		

1.	Description / Specification	
1.1	Description	ELISA Microplate, 96 well, half area well profile, solid bottom, alphanumeric well coding 675001: Microlon 200, medium binding 675061: Microlon 600, high binding
1.2	Dimensions	See Customer Drawing
1.3	Volume	Total volume: 199 µl (mathematically calculated) Working volume: 15 - 175 µl
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for immunology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	4.800 x g: Swinging-bucket rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	4 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	40
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 21 December 2021	Date 17 January 2022	Date 17 January 2022	
Date 02.12.2014	Name S. Kaelberer	Name R. Daum	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.