


PDS No. 6591xx	PRODUCT DATA SHEET			Page 1 of 1
Revision 04	60 Well HLA / Microtest Plate (Terasaki Plate), PS, Stackable			
	Greiner Item-No. 6591xx			
Valid for Item-No.:	659180	659190		

1.	Description / Specification	
1.1	Description	PS HLA / Microtest Plate (Terasaki Plate), 60 Well, conical well profile, flat well bottom, alphanumeric well coding, with lid, stackable, physical surface treatment
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	10 µl
1.4	Material / Resin	Plate and lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Plate and lid: clear
1.6	Sterilization	No
1.7	Quality Control	- Raw Material-Control: physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for cell culture to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	1 year
2.7	Other Information	-

3.	Packaging	659180	659190
3.1	Pieces / Bag	10 / bag	150 / folding carton
3.2	Pieces / Box	200	1200
3.3	Lot-No.	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 14 February 2022	Date 4 October 2022	Date 5 October 2022	
Date 02.12.2014	Name S. Kaelberer	Name P. Wachter	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.