


| | | | | | |
|------------------------------------|---------------------------------------------------------------------------------------------------|---------------|---------------|---------------|-------------------------------------------------------------------------------------|
| PDS No. 655936 65594x 655956 | PRODUCT DATA SHEET | | | | Page 1 of 1 |
| Revision 06 | 96 Well Cell Culture Microplate CELLCOAT [®] , PS, μClear [®] , Chimney Well | | | |  |
| | Greiner Item-No. 655936 / 65594x / 655956 | | | | |
| Valid for Item-No.: | 655936 | 655944 | 655946 | 655948 | 655956 |

| 1. | Description / Specification | |
|-----|-----------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1.1 | Description | PS plate, 96 well, μClear [®] , F-bottom (flat), chimney well, alphanumeric well coding, single position lid with condensation rings (high profile) 655936: Poly-L-Lysine protein coating 655944, -946, -948: Poly-D-Lysine protein coating 655956: Collagen Type I protein coating |
| 1.2 | Dimensions | See Customer Drawing Foil: 190μm ± 10% |
| 1.3 | Volume per well | Total volume: 392 μl (mathematically calculated) Working volume: 25-340 μl Growth area: 34 mm ² |
| 1.4 | Material / Resin | <u>Plate</u> : 655936: PS (Polystyrene) coated with Poly-L-Lysine 655944, -946, -948: PS (Polystyrene) coated with Poly-D-Lysine 655956: PS (Polystyrene) coated with Collagen Type I <u>Lid</u> : PS (Polystyrene) The materials for manufacturing are free of heavy metals |
| 1.5 | Colour | 655936, -946, -948, -956: plate black, bottom clear 655944: plate white, bottom clear 655936, -944, -946, -948, -956: <u>Lid</u> : clear |
| 1.6 | Sterilization | No (aseptic) |
| 1.7 | Quality Control | - <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification |
| 1.8 | Intended Use | General laboratory product for cell culture to be used by qualified personnel in a laboratory environment. |
| 1.9 | Other Information | For single use only |

| 2. | Features | |
|-----|--------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 2.1 | Basic features | - |
| 2.2 | Temperature range | Room temperature |
| 2.3 | Autoclavability | No |
| 2.4 | Centrifugation, max. RCF | 4800 x g: swinging-bucket rotor |
| 2.5 | Chemical Resistance | See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html (Only concerning the standard plate without coating) |
| 2.6 | Shelf life | 655944, -946, -948: 18 months 655936, -956: 24 months |
| 2.7 | Other Information | - |

| 3. | Packaging | 655936, -944, -946, -956 | 655948 |
|-----|-------------------|------------------------------------------------------------------------|------------------------------------------------------------------------|
| 3.1 | Pieces / Bag | 5 | 20 |
| 3.2 | Pieces / Box | 20 | 120 |
| 3.3 | Lot-No. | E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.) | E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.) |
| 3.4 | Other Information | - | - |

| 4. | Other Information |
|----|-------------------|
| | - |

Data Sheet subject to change without notice!

| Prior Issue | Drawn | Approved | Released | CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved. |
|--------------------|-------------------------|---------------------|---------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Revision 05 | Date 9 February 2022 | Date 27 May 2022 | Date 30 May 2022 | |
| Date 08.05.2019 | Name S. Kaelberer | Name P. Wachter | Name A. Illig | |

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.