PDS No. 65580x	PRODUCT DATA SHEET			Page 1 of 1	
Revision 07	96 Well Microplate, UV-Star®, Chimney Well			6	
	Greiner Item-No. 65580x			greiner	
Valid for Item-No.:	655801	655809			

1.	Description / Specification		
1.1	Description	UV-Star® Microplate, 96 well, clear film F-bottom (flat), alphanumeric well coding, chimney well	
1.2	Dimensions	See Customer Drawing Foil: 135 μm (± 10 μm)	
1.3	Volume	Total volume: 392 µl (mathematically calculated) Working volume: 25 - 340 µl	
1.4	Material / Resin	Plate and foil: Cycloolefine, free of heavy metal	
1.5	Colour	655801: <u>plate</u> ; clear 655809: <u>plate</u> ; black	
1.6	Sterilisation	No	
1.7	Quality Control	Raw Material-Control: physical testing Product-Control: testing of attributive and variable characteristics in accordance with the valid specification	
1.8	Intended Use	General laboratory product for the processing and storage of samples to be used by qualified personnel in a laboratory environment.	
1.9	Other Information	For single use only	

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens
2.2	Temperature range	For application: -20°C to +40°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	4800 x g: swinging-bucket rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	N/A
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	40
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information		
	-		

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date	Date	document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
06	8 February 2022	27 May 2022	30 May 2022	
Date	Name	Name	Name	
02.12.2014	S. Kaelberer	P. Wachter	A. Illig	