PDS No.976501		PRODUCT DATA SHEETRack for Cryo.s Biobanking Tubes 300 μl,Non-Sterile				Page 1 of 1	
Revision 04	Rack for C	ryo.s Biobankiı Non-Steril	ng Tubes 300 le	μl,			greiner
	Greiner Item-No. 976501					BIO-ONE	
Valid for Item-No.:	976501						

1.	Description / Specification	
1.1	Description	 Rack for 300 µl Cryo.s Biobanking Tubes, stackable, with a footprint that corresponds to the microplate standard of the American National Standards Institute (ANSI-Standard), capacity for 96 Cryo.s Biobanking Tubes 300 µl, rack bottom with scanning windows for the Datamatrix , rotation stoppers, snap-in lid, non-sterile. 2D Code and linear barcode 128 (G9xxxxxxx), pre-produced 2D codes are featured with a symbol size of 14x14. <u>Suitable tubes</u>: Cat No.: 131202: capped 300 µl Cryo.s, non-sterile 131263: capped 300 µl Cryo.s, sterile
		131263-2D1: capped 300 µl Cryo.s, sterile, customised coding
1.2	Dimensions	See Customer Drawing
1.3	Volume	N/A
1.4	Material / Resin	Rack / lid: PC (Polycarbonate)
1.5	Colour	Rack: black Lid: transparent
1.6	Sterilization	No
1.7	Quality Control	Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	<u>Cryo.s</u> : for storage of tissue, cells, fungi, bacteria, spores, cellular extracts or body fluids at ultra-low temperature for research and development as well as diagnostic purposes. Tubes must not be stored in the liquid phase of liquid nitrogen, but only in the gas phase above. Tubes are not intended for any application in the context of reproductive medicine.
1.9	Other Information	-

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: -196°C to 37°C
2.3	Autoclavability	N/A
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	<u>Rack / lid</u> : resistant to \rightarrow acetic acid (1 %), methanol, ethanol (96%), isopropanol, sulphuric acid (4.9%), DMSO
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	5
3.2	Pieces / Box	10
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
4.1	For Laboratory Use Only

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date		document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This
03	31 March 2023	11 July 2023		document may not be reproduced for any
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,
25.02.2021	S. Kaelberer	Dr. S. Mühlfriedel		and copyright are reserved.

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.