


PDS No. 6123xx	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 07	Serum Pipette			
	Greiner Item-No. 6123xx			
Valid for Item-No.:	<b>612301</b>	<b>612361</b> (sterile)	<b>612362</b> (sterile)	

1.	Description / Specification	
1.1	Description	Serum pipette with graduation (scale 0.5 – 2.5 ml) 612301: standard 612361, -362: sterile
1.2	Dimensions	See Customer Drawing
1.3	Volume	Working volume: 2.5 ml
1.4	Material / Resin	LDPE (Low Density Polyethylene)
1.5	Colour	Translucent
1.6	Sterilization	612361, -362: SAL 10 <sup>-3</sup> 612301: no
1.7	Quality Control	<u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for pipetting of liquids to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application -50°C to +80°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	612361, -362: 5 years . 612301: -
2.7	Other Information	-

3.	Packaging	612301	612361	612362
3.1	Pieces / Bag	500	1	25
3.2	Pieces / Box	1.500	800	1.000
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-	-	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 06	Date 19 August 2021	Date 24 August 2021	Date 31 August 2021	
Date 20.04.2015	Name S. Kaelberer	Name J. Gaiser	Name A. Illig	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.