


PDS No. 762001/ 762061	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 02	ELISA-Strip, 1x8 Well, PS, F-bottom			
	Greiner Item-No. 762001/ 762061			
Valid for Item-No.:	<b>762001</b>	<b>762061</b>		

1.	Description / Specification	
1.1	Description	ELISA-Strip, 1 x F8, PS, solid F-bottom (flat) 762001: medium binding 762061: high binding
1.2	Dimensions	See Customer Drawing
1.3	Volume per well	Total volume: 388 µl Working volume: 20-350 µl
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	No
1.7	Quality Control	- <u>Raw Material-Control</u> : physical and immunological testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for immunology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens
2.2	Temperature range	For application: -20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	4 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	100
3.2	Pieces / Box	400
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 01	Date 21 December 2021	Date 17 January 2022	Date 17 January 2022	
Date 18.05.2015	Name S. Kaelberer	Name R. Daum	Name A. Illig	

**DISCLAIMER:** The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.