


PDS No. 37327x/ 373281	PRODUCT DATA SHEET				Page 1 of 1
Revision 05	Sapphire PCR 8-Cap Strips, with Domed Lid				
	Greiner Item-No. 37327x / 373281				
Valid for Item-No.:	373270	373273	373274	373275	373276
	373277	373281			

1.	Description / Specification	
1.1	Description	Sapphire PCR 8-cap strips with domed lid for PCR. Compatible with all Sapphire 96 Well PCR microplates and Sapphire 8-tube strips (6732xx).
1.2	Dimensions	See Customer Drawing
1.3	Volume	N/A
1.4	Material / Resin	PP (Polypropylene), free of heavy metal
1.5	Colour	373270: clear 373273: red 373274: blue 373275: green 373276: yellow 373277: violet 373281: assorted (red, blue, green, yellow, violet)
1.6	Sterilization	No
1.7	Quality Control	<u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for molecular biology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens
2.2	Temperature range	For application: -20°C to +105°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	125
3.2	Pieces / Case	1.250
3.3	Lot-No.	5-digit 12345 (digit 1 & 2 identify the year and 3 & 4 identify the week)
3.4	Other Information	Packaging labelled with product symbols for easier identification and warehousing

4.	Other Information	
		-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 20 August 2021	Date 16 November 2021	Date 17 November 2021	
Date 01.07.2016	Name S. Kaelberer	Name Dr. M. Langbein	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.