


PDS No. 65589x	PRODUCT DATA SHEET		Page 1 of 1
Revision 07	96 Well SensoPlate, PS, Glass Bottom		
	Greiner Item-No. 65589x		
Valid for Item-No.:	655892 (sterile)	655896 (sterile)	

1.	Description / Specification	
1.1	Description	PS Microplate, 96 well, F-glass bottom (flat), chimney well, alphanumeric well coding, sterile 655892: with single position lid, condensation rings, high profile 655896: without lid
1.2	Dimensions	See Customer Drawing Glass bottom: 175 µm (+/- 20 µm)
1.3	Volume per well	Total volume: 392 µl (mathematically calculated) Working volume: 25 - 340 µl
1.4	Material / Resin	Plate: PS (Polystyrene), free of heavy metal Glass bottom: clear borosilicate Lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Plate: black Glass bottom: clear Lid: clear
1.6	Sterilization	SAL 10 ⁻³
1.7	Quality Control	- Raw Material-Control: physical testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for the processing and storage of samples to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	Adhesive: compatible with cell culture media, low autofluorescence
2.2	Temperature range	For application: + 4°C to +37°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Shelf life	2 years
2.6	Other Information	-

3.	Packaging	
3.1	Pieces / Air Cushion Bag	1
3.2	Pieces / Box	16
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information	
		-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 06	Date 9 February 2022	Date 27 May 2022	Date 30 May 2022	
Date 03.02.2016	Name S. Kaelberer	Name P. Wachter	Name A. Illig	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.