


PDS No. 668102	PRODUCT DATA SHEET	Page 1 of 1
Revision 04	Analyser Cup, PS, 25 ml with Cap	
	Greiner Item-No. 668102	

1.	Description / Specification	
1.1	Description	PS Analyser cup, flat bottom shape, 25 ml, with cap enclosed separately. For Coulter® and Hycel® Cellcounter-systems, high precision design for use with robotic analytical systems
1.2	Dimensions	See Customer Drawing
1.3	Volume	25 ml (Total volume)
1.4	Material / Resin	Cup: PS (Polystyrene), free of heavy metal Cap: LDPE (Low-Density Polyethylene), free of heavy metal
1.5	Colour	Cup: Clear Cap: Translucent
1.6	Sterilization	No
1.7	Quality Control	Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for sample analysis to be used by qualified personnel in a laboratory environment.
1.9	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	For application: +4°C to +30°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	N/A
2.7	Other Information	-

3.	Packaging	Analyser Cup	Cap
3.1	Pieces / Bag	250	250
3.2	Pieces / Box	1250	1250
3.3	Lot-No.	E JJ MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	
3.4	Other Information	-	

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 21 October 2024	Date 22 October 2024	Date 22 October 2024	
Date 22.04.2015	Name S. Kaelberer	Name Dr. T. Schreiber	Name Dr. C.-K. Chai	

DISCLAIMER: The description of a certain product can only be considered as a guidance because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.