PDS No.131263/ 131263-2D1	PRODUCT DATA SHEET	Page 1 of 1
Revision 07	Capped Cryo.s Biobanking Tubes 300 µl (Datamatrix Coded), Sterile	greiner
	Greiner Item-No. 131263 / 131263-2D1	BIO-ONE
Valid for Item-No.:	131263 (sterile) 131263-2D1 (sterile)	

1.	Description / Specification			
internal thread), without bottom (2D Code (G8) symbol size of 14x14)				
		(Subtoffillood Soquefillood, Suit. 140.1 of 1666), Stefillo		
		Suitable racks: Cat No.: 976501		
1.2	Dimensions	See Customer Drawing		
1.3	Volume	Working volume: 235 µl		
1.4	Material / Resin	<u>Tube / cap</u> : PP (Polypropylene), medical grade and USP class VI certified <u>Gasket</u> : Silicone, medical grade		
1.5	Colour	Tube / cap: natural Gasket (cap): blue		
1.6	Sterilization	SAL 10 ⁻⁶		
1.7	Quality Control	Raw Material-Control: physical testing Product-Control: testing of attributive and variable characteristics in accordance with the valid specification		
1.8	Intended Use	For storage of tissue, cells, fungi, bacteria, spores, cellular extracts or body fluids at ultra-low temperature for research and development as well as diagnostic purposes*. Tubes must not be stored in the liquid phase of liquid nitrogen, but only in the gas phase above. Tubes are not intended for any application in the context of reproductive medicine. *Intended use for USA: Research use only		
1.9	Other Information	For single use only		

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
		Contents non-cytotoxic
2.2	Temperature range	For application: -196°C to 37°C
2.3	Autoclavability	N/A
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	Tube / material: See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html Tube / datamatrix: resistant to → acetic acid (1 %), hydrochloric acid (25 %), sulphuric acid (4.9%), ethanol, methanol, isopropanol, DMSO
2.6	Shelf life	5 years
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	480
3.2	Pieces / Box	960
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information
4.1	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this
Revision	Date	Date	Date	document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This
06	6 June 2024	5 August 2024	7 August 2024	document may not be reproduced for any
Date	Name	Name	Name	reason without written permission from Greiner Bio-One GmbH. All rights of design, invention,
13.09.2023	S. Kaelberer	Dr. S. Mühlfriedel	Dr. CK. Chai	and copyright are reserved.